Show the calf while being activated in an MR Scanner









Highlights

High standards

Lode is a socially and environmentally responsible company. All Lode products are RoHS/WEE compliant and Lode is ISO 9001:2003, ISO 13485:2008 and FDA 510K certified. All medical products comply to MDD 93/42/EEC, incl. IEC 60601-1.

Unique ergometer for MRI

With this unique ergometer it is possible to perform an exercise test in an MRI. An ergometer on the tabletop of an MRI ensures minimal time stress induction and imaging without coursing artifacts on the imaging.

Tesla independent

The choice of materials and the special design makes that the Lode MRI ergometer is useable for a 1,5 and a 3 Tesla MRI without giving artifacts on the imaging.

Compatible with various MRIs at 1.5 and 3 Tesla

Compatible with

- Philips
- Siemens
- GE

Exercise instead of medicine

When a test subject is able to do exercise, it is always recommended above pharmacologic stress. It allows objective measurement in either level of cardiac conditioning and/or level of cardiac work. It is safe and perfectly reproducible.







Show the calf while being activated in an MR Scanner

MR Ergometer for spectroscopy of the calf (m. Tibialis anterior and m. Gastrocnemius). The workload is adjustable up to 100 watt. The zero load is <5 watt at 25 rpm. The MR ergometer is an ergometer for use during MR studies. The MR ergometers workload is controlled with an electronical braking principle especially designed for use in a MR environment. The moment of inertia is 8,4 kgm2. The MR ergometer is standard supplied with a control unit and power unit. The standard control unit offers the possibility to read out various parameters like workload, rpm, torque, timer and distance. The power unit is completed with a safety cable for wallfixation. The MR ergometer can be used for MR scanners up to 3 Tesla

Features



Siemens MRI compatibility

The MRI Ergometer can be used in combination with the following Siemens MRI scanners Magnetom Harmony, Symphony, Sonata, Espree, Avanto.



Additional features with PCU

Besides the possibility to program 24 protocols easily, this control unit offers the following features:

- better monitoring because of the additional and larger display
- a perfect combination with BPM
- possibility to measure SpO2



Philips MRI compatible

The ergometer is compatible with various Philips MRI devices like Philips Intera CV, Achieva 1.5 Tesla & 3.0 Tesla



GE MRI compatible

The ergometer is compatible with various GE MRI devices like Horizon, Twin Speed



Low noise

Due to accurate manufacturing and the careful choice of materials the product has an extremely low noise level.



Accurate over a long period of time

The Lode ergometers are supplied with an electro-magnetic braking mechanism of Lanooy (eddy current). The biggest advantage of this braking system compared to a friction braking system is the absolute accuracy and the accuracy over time. Moreover, friction braking systems have more wearing parts.



Small adjustment steps

The workload of the Lode ergometers is adjustable in steps of only 1 watt. Depending on your wishes, the test operator or the test subject can adjust the workload. The steps of 1 watt are possible in the manual mode as well as within protocols.



Service friendly ergometer

Lode ergometers are very service friendly. In general, total costs for spare parts are so low that they are negligible. Furthermore, most options are so easy to install and firmware is so easy to update that labor costs are minimal. Moreover, the ergometer can be cleaned easily.



Up till 3 Tesla

Virtually no interferentie up till 3 Tesla through smart constructions and material use





Show the calf while being activated in an MR Scanner



A unique ergometer

The Lode MRI ergometers are designed to produce physical stress within an MRI device. The MRI ergometer can be used for cardiac examinations, cardiac research, spectroscopy and other examinations and research.

For cardiac MRI examinations, the MRI ergometer can be produced with a pedal (circular) or push/pull exercise movement.

For spectroscopy MR examinations there is an ergometer available with up/down movement for the upper leg and an ankle MRI ergometer for the calf muscles.

The MRI ergometers are compatible for the most types of MRI scanners of Siemens, Philips and GE. The choice of materials and the special design makes that the Lode MRI ergometer can be used for 1,5 and a 3 Tesla MRI without giving artifacts on the imaging.

Our MRI ergometer with its low start-up load enables exercise. When a test subject is able to do exercise, this is always recommended above pharmacologic stress. It allows objective measurement of improvement in either level of cardiac conditioning and/or level of cardiac work. It is safe and what is very important is perfectly reproducible.





Show the calf while being activated in an MR Scanner

MR Ergometer Dorsal Ankle Flexion can a.o be extended with the following options:







Show the calf while being activated in an MR Scanner

Specifications

Workload

Workload		
Minimum load	5 W	
Maximum peak load	100 W	
Minimum load increments	1 W	
Maximum continuous load	100 W	
Hyperbolic workload control	~	
Linear workload control	~	
Fixed torque workload control	~	
Maximum rpm independent constant load	60 rpm	
Minimum rpm independent constant load	5 rpm	
Electromagnetic "eddy current" braking system	~	
Dynamic calibration	~	
Accuracy		
Workload accuracy below 100 W	3 W	
Workload accuracy from 100 to 500 W	3 %	
User Interface		
Manual operation mode	~	
Analog operation mode	~	
Terminal operation mode	~	
External control unit	~	
Selfdesigned protocol operation mode	~	
Connectivity		
Analog connector	~	
Dimensions		
Product length (cm)	135 cm	53.1 inch
Product width (cm)	50 cm	19.7 inch
Product height	50 cm	19.7 inch
Product weight	47 kg	103.6 lbs
Power requirements		
115 V AC 50/60 Hz (130 VA)	~	
230 V AC 50/60 Hz (130 VA)	~	
Standards & Safety		
IEC 60601-1:2005	~	
ISO 13485:2003 compliant	~	
ISO 9001:2008 compliant	~	
Certification		
CE class Im according to MDD93/42/EEC	~	
CTüVus according to NRTL	~	
CB according to IECEE CB	~	

Order info

Partnumber: 937904

^{*}Specifications are subject to change without notice.



Distributed by

Costa Rica

Costa Rica

Zernikepark 16 9747 AN Groningen The Netherlands Tel: +31 50 5712811 Fax: +31 50 5716746 E-mail: ask@lode.nl Internet: www.lode.nl

Lode B.V.